

### CERTIFICATE OF TEST

Test certificate n°:  
"IBE-BVI" – P-15.176/PP60JJ & PO60JJ"

Date certificate: 24 July 2015

Test Sample: Manufacturer references:  
Sharps container Type: PP60JJ/PO60JJ  
Material : Polypropylene  
Dimensions: (400x330x680) mm  
Weight : PP 1980 g / PO 2100 g  
Real capacity : 53 litres



Owner: **PLAZUR**  
Avenue du 14 juillet, 60J  
F - 21300 CHENÔVE  
FRANCE

Manufacturer: **Container**  
**Bichat SAS**  
20, Route de Bourg  
F - 01250 Villereversure

**Lid**  
**BOUILLARD SAS**  
38 route de Louhans  
F - 71370 ST GERMAIN DU PLAIN

Test: According to EN ISO 23907 (October 2012): Sharps injury protection – Requirements and test methods – Sharps containers and NF X 30-511 (April 2015): Containers for medical care waste – Additional and/or alternative characteristics and requirements for sharps containers

Test Results: The present materials comply with the European Norm EN ISO 23907 (October 2012) and French norm NF X 30-511 (April 2015).  
The complete description of the presented material, the results and findings of the tests are given in the report **PBO/pbo/P-14.079 of 03-Apr-2014**, **PBO/pbo/P-14.131 of 24-Apr-2014** and **PBO/pbo/P-15.176-C of 24-Jul-2015** which is inextricable to this certificate

The Belgium Packaging Institute Ltd, Z.I. Research Park 280, B-1731 Zellik, certifies the here above mentioned statements. Any change of the presented material or production method involves a new series of conformity tests on the modified packages to be performed by the Belgium Packaging Institute.

**The certificate and report are valid for an unlimited period as far as the production method remains identical and the packaging remains within the specifications of the tested samples.**

Quality Manager



Filip Van Impe



Analyst



Philippe Bourdon

page 1/1

*This certificate of test and its report shall not be reproduced except in full, without written approval of the B.V.I./I.B.E.*